



CT/CBCT SCAN

1. Use a slice thickness of 0.4–0.5 mm
2. Make sure the patient doesn't move
3. The occlusal surface of the arch of interest has to be horizontal
4. The field of view must cover the arch of interest and 5 mm of opposing arch
5. Make sure patient's teeth are separated
6. Separate cheeks and lips from gingiva

